



BI
Biological Industries

EZ Lympho-Sep™ Lymphocyte Separation Tubes

For density gradient
separation of lymphocytes

Cat. No.: 01-899-U02
01-899-U04
01-899-U10
01-899-U16

Store at: 4-25°C

Instructions for Use

Product Description

Density gradient centrifugation of diluted whole blood on a polysucrose - sodium metrizoate medium is the method of choice for isolation of lymphocytes. The success of the procedure, i.e. the recovery of viable lymphocytes with the lowest proportion of contaminating granulocytes and erythrocytes, depends to a large extent on the careful layering of the blood sample onto the polysucrose - sodium metrizoate and the maintenance of a sharp interface between the two solutions prior to centrifugation. To this end EZ Lympho-Sep™ was developed. The EZ Lympho-Sep™ system allows the blood sample to be poured directly into the centrifuge tube with no special precautions required to prevent disruption of the polysucrose - sodium metrizoate layer. Thus, a large number of samples may be handled at the same time. The mechanism also reduces the length of centrifugation time required for separation of the lymphocytes.

Precaution and Disclaimer

For research use only!

Principle

EZ Lympho-Sep™ products are sterile plastic centrifuge tubes containing a solution of 5.6% polysucrose and 9.6% sodium metrizoate (density 1.077 g/ml, osmolality 280 mOsm), which is sequestered in the bottoms of the tubes by one-way plastic inserts. Blood poured onto the insert does not disrupt the surface of the polysucrose - sodium metrizoate layer. During centrifugation the insert opens to permit mixing of the blood and polysucrose - sodium metrizoate. Aggregated red blood cells and granulocytes are sedimented to the bottom of the tube while the white blood cells migrate to the interface between the plasma and polysucrose - sodium metrizoate layers. They are readily identified as a discrete white band located above the plastic insert.

Instructions for Use

1. EZ Lympho-Sep™ products are sterile and ready-for-use. Open only under aseptic conditions.
2. Best results are obtained when all steps are performed at 18-20°C.
3. Use anticoagulant treated or defibrinated blood. Blood may be diluted with an equal volume of sterile saline or other sterile isotonic buffer or may be used undiluted. Add diluted or undiluted blood, according to Table 1. directly to the tube. Cap and centrifuge (18-20°C) 1000 x g for 20 min. Procedures carried out at lower temperature may require longer centrifugation.

Table 1.

| Cat. No. | Diluted Blood (1:1) | Undiluted Blood |
|------------|---------------------|-----------------|
| 01-899-U02 | 4 - 8 ml | 2 - 4 ml |
| 01-899-U04 | 4-11 ml | 2 - 5.5 ml |
| 01-899-U10 | 20 -35 ml | 10 -17.5 ml |
| 01-899-U16 | not applicable | 18.5-25 ml |

Erythrocytes, dead cells and PMNs (polymorph nuclear leukocytes or granulocytes) are found at the bottom of the tube. The EZ Lympho-Sep™ insert separates the lymphocyte interface from the pellet of packed erythrocytes.


4. Remove the platelet-rich plasma and discard it.
5. Remove the mononuclear layer with the aid of a pipette. Alternatively, the entire contents of the tube above the plastic insert may be removed by decanting the solution.

Storage

Store at 4-25°C out of direct light. Deterioration of the polysucrose - sodium metrizoate is indicated by the appearance of a distinct yellow color or particulate material in the clear solution.

Note

Under normal shipping conditions the polysucrose - sodium metrizoate remains below the insert. Improper handling may cause some leakage to the top of the insert. In this case, centrifuge the tube for 1 minute at 400 x g to displace the liquid to the bottom of the tube before use. Plastic centrifuge tubes are somewhat brittle. Care should be taken to avoid dropping the tubes or striking them against hard surfaces.



BI
Biological Industries
Culture of Excellence

Biological Industries Israel Beit Haemek Ltd.
Kibbutz Beit Haemek 25115, Israel
T.+972.4.9960595 F.+972.4.9968896
Email: info@bioind.com
www.bioind.com



cGMP
Manufacturing
Facility